

Dear KHANGAN/NANDHINIS.LO

Sanction No. SS4/R/111091 dated 13/03/2025 has been generated in reference to Service order No. 371351 dated 04/03/2025.

Area : SS4 COIMBATORE

Model : 41103015SD00028/GODREJ VACCINE REF 225L AC

Replacement Product Code : 41103015SD00027 MR GBR 225L AE WHITE

Product Category : REFRIGERATOR

Defective Product Code : 41103015SD00028 GODREJ VACCINE REF 225L AC

Cabinet No(For Appl) and SO/RPL/Invoice No(For Others) : 210300028MR00356

Invoice No : 66447

Invoice Date : 09/08/2021

DOP : 09/08/2021

DEALER / CUSTOMER : CUSTOMER

Damage/Defective : Defective

Nature Of Problem : Customer_Expectations_-_Not_allowing_Repairs

Job Code Desc : WATER LEAKAGE#LEAK FROM INNER CABINET#APPLIANCE TO BE REPLACED

PRA submitted by : ABISHEGK.R

Customer / Dealer Code : 40000E

Customer / Dealer Name : GOVT PRIMARY HEALTH CENTER GOVT PRIMARY HEALTH CEN

Address : GOVT PRIMARY HEALTH CENTER GOVT PRIMARY HEALTH CENTER GOVT PRIMARY
HEALTH CENTER; VANGAL KARUR

City: VANGAL

Telephone Number : 9629874211

Mobile Number : 7338843705

Observation text: CUSTOMER CASE-ILR- WATER TANK LEAKAGE CASE MR.PRATIK AGARWAL APPROVAL ATTACHED PLEASE APPROVE THE SAME FOR REPLACEMENT SIR

Approval remark: ILR CABINET LEAK CASE HO APPROVAL IS ATTACHED

Sanction text: Sanctioned as per HO approval mail Price Diff Rs.-88188.56/-
Waived off APRV By Mr. Pratik Agarwal

Kindly take necessary action.

ADHI

Note to commercial team -

- In case of any deviation (to actual machine) to above details , than above sanction no. stands void.

- If any alteration required in above details than this document needs to be regenerated (pdf format) through smartnet system. Contact respective service executive.